

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	SL POWER ELECTRONICS CORP BLDG A 6050 KING DR VENTURA CA 93003 USA
Name and address of the manufacturer	SL POWER ELECTRONICS CORP BLDG A 6050 KING DR VENTURA CA 93003 USA
Name and address of the factory	SL XIANGHE POWER ELECTRONICS CORP No.B-02-03, North side of Landscape Ave, Qibu Distr Environmental Industrial Park Xianghe,Hebei 065400 China
Note: When more than one factory, please report on page 2	<input checked="" type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	Input: 100-240Vac, 50-60Hz, 4.0A <input checked="" type="checkbox"/> Additional Information on page 2
Trademark (if any)	N/A
Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	SLB300SXXYZZ <input checked="" type="checkbox"/> Additional Information on page 2
Additional information (if necessary may also be reported on page 2)	The risk management requirements of the standard were not addressed. For Class I. <input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E511446-D1022-1/A0/C0-CB issued on 2021-06-30
This CB Test Certificate is issued by the National Certification Body	



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2021-07-20

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-38489-UL

Factories:

INDUSTRIAS S L S A DE C V
CIRCUITO SIGLO XXI 2055 COL PARQUE INDUSTRIAL EX-XXI 21254 MEXICALI BC
Mexico

Additional Model(s):

Series: SLB300SXXYZZ, Where XX represents the output voltage which may be any number from 12 to 56. Y can be K (for Class I); ZZ can be any number between 00-99, blank or any letter from AA to ZZ, only for market purpose, not influence safety function.

Additional Ratings:

Output: 12-56Vdc, 22.5-5.36A, 300W max.
Standard models output rating are listed as below:
For SLB300S12YZZ: 12Vdc/22.5A, Max. 270W
For SLB300S24YZZ: 24Vdc/12.5A, Max. 300W
For SLB300S48YZZ: 48Vdc/6.25A, Max. 300W
For SLB300S56YZZ: 56Vdc/5.36A, Max. 300W

Additionally evaluated to:

EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014
National Differences specified in the CB Test Report.

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Signature:

Jolanta M. Wroblewska

Date: 2021-07-20